

Band-Lok® & Band-Lok® Blue Dual Cure Compomer Band Cement

INTENDED USE: Band-Lok® is a two paste, dual-cure adhesive intended for use as a band and appliance cement.

INSTRUCTIONS FOR USE:

Step 1. Prophyl, rinse, dry and isolate tooth to be banded. Acid etching of the enamel is not required for cementing bands, however, it will increase strength required for high stress appliances.

Step 2. On a mixing pad place equal parts of Band-Lok® Pastes A and B (A 1 inch strip of each part will provide enough cement for 4 bands). If paste is not going to be mixed immediately, shield from light.

Step 3. Mix Pastes A and B thoroughly for 10 seconds and place in band. Seat band and clean off excess flash.

Step 5. At this point you have three curing options:

Light cure the cement from occlusal for 10 seconds. Cement can be exposed to saliva at this point. Final cure will occur in 5 minutes.


Light cure the cement from the occlusal for 30 seconds. Cement is now completely cured and can be exposed to forces immediately.

Allow the cement to chemically cure on its own; complete polymerization will occur in 10 minutes.

IMPORTANT: Band-Lok® will adhere to metal and stainless-steel instruments! Before using, spray instruments with a light coat of PAM® or oil-free silicone before use.

SELF MIXING DUAL-SYRINGE DELIVERY SYSTEM

1. Remove cap and place mixing tip on syringe
2. Syringe material into band/crown or appliance using mixing tip.
3. Replace the cap on the syringe after use.



! **PRECAUTIONS:** Unpolymerized resins may cause skin sensation in susceptible persons. Wash hands and other exposed areas with mild soap and water before eating, drinking or smoking and when leaving work. Provide good ventilation in process area to prevent formation of vapor. Rinse eyes immediately with plenty of water. Obtain medical attention if pain, blinking or redness persist.

CAUTION: DO NOT REUSE IF SYRINGE BECOMES CONTAMINATED. AVOID CONTACT OF THE REUSABLE PARTS WITH THE PATIENTS MOUTH. USE NEW, UNCONTAMINATED GLOVES. APPLY DISPOSABLE BARRIER SLEEVES/WRAPPS BEFORE EACH PATIENT

② **Important! Self-Mixing Tips are single use devices. A new tip is required for each patient.**

⚠ Use of UV, halogen and LED curing lights can be associated with eye damage. The patient, doctor and the assistant should wear properly designed protective eyewear when using these light curing devices.



TEMPERATURE LIMITATION: Store refrigerated (2-8 Degrees C.)

To Reorder: Kits: Band-Lok® Natural Shade Kit (REF# BL), NonFluoride (REF#NFBL), Band-Lok® Blue Shade Kit (REF # BLB) or NonFluoride (REF# NFBLB)

Components: Band-Lok® Natural Shade Part A Syringe (REF # BLSA) or NonFluoride (REF#NFBSA), Band-Lok® Blue Shade Part A Syringe (REF# BLSAB) or NonFluoride (REF# NFBSAB), Band-Lok® Part B Syringe (REF# BLSB), Band-Lok® in Self Mixing Syringes Natural Shade (REF# SMBL) or NonFluoride (REF# NFSMBL), Band-Lok® in Self Mixing Syringes Blue Shade (REF# SMBLB) or NonFluoride (REF# NFSMBLB)

Contents of Kit:

4 x 6 gm syringes Paste A, 4 x 6gm syringes Paste B, Mixing Pad and Spatulas.

WARRANTY: Reliance Orthodontic Products, Inc recognizes its responsibility to replace products if proven to be defective. Reliance Orthodontic Products, Inc does not accept liability for any damages or loss, direct or consequential, stemming from the use of or inability to use the products as described. Before using, it is the responsibility of the user to determine the suitability of the product for its intended use. The user assumes all risk.

Rx Only: Federal law restricts this device to sale by or on the order of a dental professional



Reliance Orthodontic Products, Inc.

1540 West Thorndale Ave., Itasca, IL 60143 U.S.A.
1-800-323-4348 · www.RelianceOrthodontics.com



EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands



Made in the USA

Australian Sponsor: Emergo
Australia
201 Sussex St. Darling Park
Tower II, Level 20,
Sydney, NSW 2000 Australia

Check out our new learning
platform Reliance U!
www.RelianceU.tech

Scan for more
information